

The Nawanshahr Cooperative Sugar Mills Limited, Nawanshahr (Punjab)

TENDER FORM

Sealed Tenders are invited from Manufacturers/Authorised Dealers, Stockists, reputed firms for the purchase and registration of firms in order of preference for the following.

S.No.	Description of material.	Quantity	Rate
1.	Horizontal Air Laminar Flow For Plant Tissue Culture Laboratory (With installation at Site) (Size 4x2x2 feet) Make Comech/Thermadyne/Klenzaid/Equitron/Other Working size: 1200(w) x 610 (d) x 610 (h) mm Outer size: 1250(w) x 1010 (d) x 1460 (h) mm Technical Specifications - MOC : All External Body part MS power Coated Table top : SS304 - HEPA Filter: H-13 grade,efficiency 99.99% down to 0.3 microns frame extruded aluminum, micro fiber glass, perforated grill at HEPA filter SS/MS power coated - Suction Filter at : 5 microns pre-filter washable,extruded aluminium frame, TOP/Bottom synthetic non woven - Air Velocity : 0.35 – 0.65 m/s - Air Cleanliness : Class 100 - Power Supply : 220V/50Hz Single Phase - Motor : Single phase with backward curve blower - Pressure gauge: 01 nos across HEPA Filter - Steripot : 150x155 x225 mm (LxWxH), Preset Temp. Range 50° C-300° C, MOC- outer MS Inner SS-304 - Noise level : >65-70 dB - Standard Fitting : LED Light Germicidal UV light Sash (Front Foldable transparent door : Gas Cock- 01 nos ON/OFF Switches (Light, UV, Blower)	01 Nos	

Note:-The quantity of material may increase or decrease at the time of order.

TERMS AND CONDITIONS:-

- Last date for receipt of tender is 05.11.2024 upto 11:00 AM which will be opened and negotiation will be held on 05.11.2024 at 02:00 PM with the lowest quote only, if required.
- Material offered should be strictly as per our specifications and in the Tender Form Issued by the Mills only.
- Minimum delivery period must be mentioned.
- One percent Penalty per week of delay subject to maximum of 5% will be imposed.
- GST extra as applicable.
- 80% payment will be made after receipt of material and balance 20% after quality verification of material within 20 days.
- Rates should be FOR Mills, GST if applicable should be mentioned Separately in percentage.
- The envelope of tender should be superscribed as Tender for “Spares for Autoclave and Distillation “PERICLAVE” make”.
- Jurisdiction : Any dispute there items shall be subject to jurisdiction of Nawanshahr.
- The undersigned reserves the right to reject/accept any or all tenders without assigning reasons.

General Manager

CCDO